



REVISIONS				
ZONE	REV	DESCRIPTION	DATE	APPROVED

Copper tube is 25-1/8" long.
The end bends are done after
tube is inserted into the
power supply module and the
epoxy has cured.

1/8" Copper Tubing

PPD/ETT/DO Electronics Group David L. Huffman	FERMI NATIONAL ACCELERATOR LABORATORY UNITED STATES DEPARTMENT OF ENERGY		
	Calorimeter Electronics Upgrade Vicor Power Supply Mod. Cooling Tube Fixture & Assembly		
	SIZE D	FSCM NO. 	DWG NO. 3823.121.MD-330265
03/11/99	SCALE 1:1		REV
		SHEET	